

# BK MODIFICATION GUIDE

## TT Template Socket Modifications



(877) 737-8444 • fax (407) 290-2419 • custserv@orthomerica.com

Practitioner Name \_\_\_\_\_ Company Name \_\_\_\_\_

Account Number \_\_\_\_\_ Date \_\_\_\_\_

Modifications	Total Surface Weight Bearing	Specific Weight Bearing or PTB	Clinician's Preferences
PTB Bar	0	-3mm or 1/8"	
Interosseous Space	0	-2mm or 5/64"	
Medial Condylar Flare	0	-3mm or 1/8"	
Fibular Head	0	+4mm or 5/62"	
Tibial Crest	0	+4mm or 5/62"	
Anterior Distal Tibia	0	0	
Elongation MPT to Distal End	0	0	
Shortening MPT to Distal End	-6mm or 1/4"	0	
% Volumetric Reduction MPT to Distal End	4%	0%	
% Volumetric Reduction 2" below MPT	0	0	
% Volumetric Reduction 4" below MPT	0	0	
% Volumetric Reduction 6" below MPT	0	0	
% Volumetric Reduction 8" below MPT	0	0	

Trimline	TSWB	SWB or PTB	Clinician's Preferences
Anterior	40mm or 1 1/2" proximal to MPT		
Posterior A "W"	15mm or 5/8" distal to MPT	15mm or 5/8" distal to MPT	
Posterior B Slanted	Slanted, lower medially to 15mm or 5/8" distal to MPT		
Medial/Lateral	56mm or 2 1/2" proximal to MPT	75mm or 3" proximal to MPT	
Soderberg Trim lines			

Socket guide on following page.

